

EU DECLARATION OF CONFORMITY

FOR MEDICAL DEVICES AND PERSONAL PROTECTIVE EQUIPMENT

Originator: J.F ROBLES Revision: 12 Revision date: 27.02.2020 Validity date: 27.02.2025

PRODUCT	SHIELDskin™ ORANGE NITRILE™ 300
DESCRIPTION	Powder Free Extra Length Ambidextrous Non- Sterile 30 cm Nitrile Gloves
CLASSIFICATION	Medical Device Class 1 / Personal Protective Equipment (PPE) Category III (Complex Design)

SHIELD Scientific codes	Sizes
67 6251	6/XS
67 6252	7/S
67 6253	8/M
67 6254	9/L
67 6255	10/XL
67 6256	11/XXL

The manufacturer established in the Union:

SHIELD Scientific B.V.

Dr Willem Dreeslaan 1 - 6721 ND BENNEKOM - THE NETHERLANDS

declares under his/her sole responsibility that the Medical Device and PPE (product codes as mentioned above) described hereafter:

SHIELDskin™ ORANGE NITRILE™ 300

is in conformity with the provisions of Council Directive 93/42/EEC and with the national standards transposing harmonized standards EN 455-1:2000, EN 455-2:2015, EN 455-3:2015 and EN 455-4:2009. It is self-certified as a Medical Device Class 1.

is in conformity with the provisions of Regulation (EU) 2016/425 and with the harmonized standards EN ISO 374-1:2016 (as a Type B glove against reagents: J, K, P & T), EN 374-2:2014 (performance level 3, including protection against viruses), EN 16523-1:2015, EN 374-4:2013, EN ISO 374-5:2016 and EN 420:2003 + A1:2009. This device is identical to the PPE, which is the subject of EU Type Examination (Module B) certificate of conformity no. FI18/962117 issued by the Notified Body:

SGS FIMKO OY (Notified Body No: 0598) Takomotie 8, FI-00380 Helsinki, Finland

This device is subject to the procedure set out in Article VIII (Module D) of the Regulation under the surveillance of the Notified Body:

SGS FIMKO OY (Notified Body No: 0598) Takomotie 8, FI-00380 Helsinki, Finland

Signed for and behalf of SHIELD Scientific B.V

SHIELD Scientific

J.F ROBLES

General Manager

Date: 27th February 2020

Place: Bennekom

Validity of this declaration: 27th February 2020 until 27th February 2025